

OPEN ACCESS

**sensors**

ISSN 1424-8220

[www.mdpi.com/journal/sensors](http://www.mdpi.com/journal/sensors)

*Supplementary Information*

## **Monoterpenes Released from Fruit, Plant, and Vegetable Systems. *Sensors* 2014, 14, 18286–18301**

**Mohammad Asif Iqbal, Ki-Hyun Kim \* and Jeong Hyeon Ahn**

Atmospheric Environment Laboratory, Department of Civil and Environmental Engineering, Hanyang University, 222 Wangsimni-ro, Seoul 133-791, Korea;  
E-Mails: iqbaldu88@gmail.com (M.A.I.); qqq112311@gmail.com (J.H.A.)

\* Author to whom correspondence should be addressed; E-Mail: kkim61@hanyang.ac.kr or kkim61@nate.com; Tel.: +82-2-2220-2325; Fax: +82-2-2220-1945.

---

**Table S1.** Preparation procedures of liquid phase working standard (L-WS) of MTs used for ST/TD/GC analysis (by both MS and FID).

<b>A. Preparation of Liquid Phase Standard for 10 Target MTs Selected in This Study</b>													
	<b>Compounds</b>	<b>Methanol</b>	<b><math>\alpha</math>-PN</b>	<b>CMP</b>	<b><math>\beta</math>-PN</b>	<b>3-CN</b>	<b>MRC</b>	<b><math>\alpha</math>-PD</b>	<b><math>\alpha</math>-TP</b>	<b>R-LN</b>	<b><math>\gamma</math>-TP</b>	<b>p-CM</b>	<b>T</b>
<b>RGC <sup>a</sup></b>	Purity (%)		99.0	95.0	98.5	98.5	95.0	95.0	95.0	99.0	98.5	99.0	99.5
<b>PS <sup>b</sup></b>	Volume ( $\mu$ L)	8,790	1,000	1,050	1,000	1,000	1,160	1,000	1,000	1,000	1,000	1,000	1,000
	Concentration (ng/ $\mu$ L)		42,471	43,192	42,946	42,207	43,584	40,375	39,758	41,679	41,863	42,570	43,283
<b>1st WS <sup>c</sup></b>	Volume ( $\mu$ L)	19,750	250	250	250	250	250	250	250	250	250	250	250
	Concentration (ng/ $\mu$ L)		531	540	537	528	545	505	497	521	523	532	541

  

<b>B. Preparation of L-WS for DI and ST- Based Calibration</b>													
<b>Order</b>	<b>Mixing Volume (<math>\mu</math>L)</b>		<b>Concentration <sup>d</sup> (ng/<math>\mu</math>L)</b>										
	<b>1st WS</b>	<b>Methanol</b>	<b><math>\alpha</math>-PN</b>	<b>CMP</b>	<b><math>\beta</math>-PN</b>	<b>3-CN</b>	<b>MRC</b>	<b><math>\alpha</math>-PD</b>	<b><math>\alpha</math>-TP</b>	<b>R-LN</b>	<b><math>\gamma</math>-TP</b>	<b>p-CM</b>	<b>T</b>
1	14	1,486	5	5	5	5	5	5	5	5	5	5	5
2	28	1,472	10	10	10	10	10	9	9	10	10	10	10
3	56	1,444	20	20	20	20	20	19	19	19	20	20	20
4	140	1,360	50	50	50	49	51	47	46	49	49	50	50
5	280	1,220	99	101	100	98	102	94	93	97	98	99	101
6	850	650	301	306	304	299	309	286	282	295	297	302	307

<sup>a</sup> Reagent grade chemicals; <sup>b</sup> PS: Dilution of pure chemicals (reagent grade chemical) to make 20 mL solution; <sup>c</sup> 1st L-WS: Dilution of PS to make in 20 mL solution; <sup>d</sup> Injection/loading volume on ST: 1  $\mu$ L.

**Table S2.** Instrumental setup and operational conditions for the gas chromatographic analysis of MTs.

<b>A. Basic Information of Sorbent Tubes (STs) Prepared for the Absorption of MTs</b>					
Sorbent material-	Tenax TA + Carbopack B + Carbopack X	Short name		TBX	
Tube material	Quartz	Amount of each sorbent	50 mg (all in quartz)		
Sorbents separated by	Quartz wool	Purge/sweeping gas	N <sub>2</sub> (99.999 %)		
<b>B. GC/MS System: QP2010, Shimadzu, Japan</b>					<b>(b) Detector (MS)</b>
<b>(a) Oven setting</b>		Ionization mode:		EI (70 eV)	
Initial temp.:	80	°C	Ion source temp.:	200	°C
Ramping rate:	1	°C min <sup>-1</sup>	TIC scan range:	35~250	m/z
Final temp.:	90	°C	Threshold:	100	
Initial hold time	0	min	<b>(c) Column (CP-WAX column, Varian, USA)</b>		
Final hold time:	5	min	Length ( <i>l</i> )	60	m
Total run time	15	min	Internal diameter ( <i>id</i> )	0.25	mm
Carrier gas:	He	99.999%	Film thickness:	0.25	μm
<b>C. GC/FID: Varian GC, Agilent Tech, USA; Column: CP-WAX (<i>l</i>: 30 m, <i>id</i>: 0.25 mm, and film thickness: 0.25 μm)</b>					
<b>(a) Oven setting</b>		<b>(b) Detector setting</b>			
Initial temp	75	°C	Detector temp.	250	°C
Initial rate	2	°C·min <sup>-1</sup>	H <sub>2</sub> and N <sub>2</sub> flow	30	mL·min <sup>-1</sup>
Initial Hold	1	min	N <sub>2</sub> flow	30	mL·min <sup>-1</sup>
Final temp	90	°C	Air flow	300	mL·min <sup>-1</sup>
Final Hold	1.5	min	Carrier gas N <sub>2</sub>	1	mL·min <sup>-1</sup>
Total time	10	min			
<b>D. Thermal Desorber (TD): UNITY, Markes International, Ltd., UK <sup>a</sup></b>					
Cold trap:	Carbopack C + Carbopack B				
split ratio:	1:5		trap low:	-10	°C
split flow:	20	mL·min <sup>-1</sup>	trap high:	300	°C
trap hold time:	5	min	flow path temp.	120	°C
desorption time:	10	min	Desorption temp.	300	°C

<sup>a</sup> TD setup was identical for both GC/FID and GC/MS analysis.

**Table S3.** Basic quality assurance (QA) parameters for target MTs analyzed by both FID and MS.

Order	Compounds	MDL <sup>a</sup>				RSE <sup>b</sup> (%)	
		TD/GC/MS		TD/GC/FID		MS	FID
		Mass (ng)	Concentration (ppb) <sup>c</sup>	Mass (ng)	Concentration (ppb)		
1	$\alpha$ -PN	0.24	0.09	0.38	0.14	0.50	1.17
2	CMP	0.23	0.08	0.89	0.32	1.06	0.5
3	$\beta$ -PN	0.38	0.14	0.72	0.26	3.76	2.04
4	3-CN	0.40	0.15	0.53	0.19	0.63	4.04
5	MRC	0.24	0.09	0.54	0.19	1.67	2.17
6	$\alpha$ -PD	0.25	0.09	0.72	0.26	1.66	1.10
7	$\alpha$ -TP	0.41	0.15	0.64	0.23	2.23	1.79
8	R-LN	0.50	0.18	0.62	0.22	0.83	1.43
9	$\gamma$ -TP	0.34	0.12	0.58	0.21	1.02	2.48
10	p-CM	0.29	0.11	0.6	0.22	0.85	0.66
11	T	0.56	0.30	0.43	0.23	2.12	1.24
Average MDL & RSE values for all MTs		0.33	0.12	0.62	0.22	1.42	1.74

<sup>a</sup> MDL values were obtained by conducting seven replicate analyses of 1  $\mu$ L liquid standard; (5 ng· $\mu$ L<sup>-1</sup>) loaded on to ST with the supply of 0.5 L N<sub>2</sub> purging gas; <sup>b</sup> Derived by triplicate analyses of 1  $\mu$ L of liquid standard (5 ng· $\mu$ L<sup>-1</sup>) loaded on to ST; <sup>c</sup> DL values for gas standards were calculated in ppb unit assuming the total sample volume of 0.5 L.

© 2014 by the authors; licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution license (<http://creativecommons.org/licenses/by/4.0/>).