

# **Optical fiber ball resonator sensor spectral interrogation through undersampled KLT: application to refractive index sensing and cancer biomarker biosensing**

**Daniele Tosi<sup>1,2,\*</sup>, Zhannat Ashikbayeva<sup>1</sup>, Aliya Bekmurzayeva<sup>1,2</sup>, Zhuldyz Myrkhiyeva<sup>1</sup>, Aida Rakhimbekova<sup>1</sup>, Takhmina Ayupova<sup>1</sup>, and Madina Shaimerdenova<sup>1</sup>**

<sup>1</sup>Nazarbayev University, School of Engineering and Digital Sciences, 010000 Nur-Sultan, Kazakhstan

<sup>2</sup>National Laboratory Astana, Laboratory of Biosensors and Bioinstruments, 010000 Nur-Sultan, Kazakhstan

\*Correspondence: daniele.tosi@nu.edu.kz;

**Table S1.** The fabrication parameters of the ball resonator using CO<sub>2</sub> laser splicer (Fujikura LZM-100). Pre-heat value and the power levels are reported in bit units, as displayed on the splicer.

Parameters	Diameters (x, y axes), μm
	<b>490-484</b>
Pre-heat (bit)	1
Absolute power (bit)	342
Relative power (bit)	90
Break-add power (bit)	100
Feeding speed (mm/sec)	0.2
Rotator speed (deg./sec)	180
Diameter adjustment (μm)	-10