

Supplementary Materials: Prevention and Treatment of Fungal Skin Infections Using Cationic Polymeric Films

Fritz Ka-Ho Ho, Albert Bolhuis and M. Begoña Delgado-Charro

Table S1. Analytical and calibration parameters for the detection of iron using ICP-OES.

Parameter	Fe
Calibration range	0.1–50 ppm
Limit of Detection (LoD)	3.57 ppm
Limit of Quantitation (LoQ)	11.92 ppm
R^2	0.99

Table S2. Analytical and calibration parameters for the detection of glucose and mannose using LCMS.

Parameter	Glucose	Mannose
Calibration range	10–200 $\mu\text{g mL}^{-1}$	10–200 $\mu\text{g mL}^{-1}$
Limit of Detection (LoD)	46.24 $\mu\text{g mL}^{-1}$	44.29 $\mu\text{g mL}^{-1}$
Limit of Quantitation (LoQ)	154.12 $\mu\text{g mL}^{-1}$	147.63 $\mu\text{g mL}^{-1}$
R^2	0.98	0.98