

Reliability and Agreement of Three Devices for Measuring Implant Stability Quotient in the Animal Ex Vivo Model

Monica Blazquez-Hinarejos ¹, Constanza Saka-Herrán ¹, Victor Diez-Alonso ¹, Raul Ayuso-Montero ^{2,3,*}, Eugenio Velasco-Ortega ⁴ and Jose Lopez-Lopez ^{3,5}

¹ Faculty of Medicine and Health Sciences (Dentistry), University of Barcelona, 08007 Barcelona, Spain; mblazquezhinarejos@gmail.com (M.B.-H.); constanzasakah@gmail.com (C.S.-H.); victor.diez.a@gmail.com (V.D.-A.)

² Department of Odontostomatology, Faculty of Medicine and Health Sciences (Dentistry), University of Barcelona, 08007 Barcelona, Spain

³ Oral Health and Masticatory System Group (Bellvitge Biomedical Research Institute) IDIBELL, University of Barcelona, 08007 Barcelona, Spain

⁴ Department of Stomatology, Faculty of Dentistry, University of Seville, 41009 Seville, Spain; evelasco@us.es

⁵ Department of Odontostomatology, Faculty of Medicine and Health Sciences (Dentistry), Barcelona University Dental Hospital, University of Barcelona, 08007 Barcelona, Spain; 18575jll@gmail.com

* Correspondence: raulayuso@ub.edu

Table S1. Normality and homogeneity of variances.

Operator / Device	Shapiro-Wilk		Device	Levene Test	
	Statistic	Significance		Statistic	Significance
MB / OSSTELL [®] Beacon parallel	0.97	0.45	OSSTELL [®] Beacon parallel	0.10	0.76
MB / OSSTELL [®] Beacon perpend.	0.98	0.79	OSSTELL [®] Beacon perpend.	0.62	0.43
MB / PENGUIN [®] parallel	0.95	0.18	PENGUIN [®] parallel	0.05	0.83
MB / PENGUIN [®] perpend.	0.97	0.61	PENGUIN [®] perpend.	0.88	0.35
MB / MEGAIQ [®] parallel	0.96	0.26	MEGAIQ [®] parallel	1.50	0.23
MB / MEGAIQ [®] perpend.	0.99	0.97	MEGAIQ [®] perpend.	1.42	0.24
RA / OSSTELL [®] Beacon parallel	0.94	0.07			
RA / OSSTELL [®] Beacon perpend.	0.94	0.10			
RA / PENGUIN [®] parallel	0.97	0.44			
RA / PENGUIN [®] perpend.	0.98	0.74			
RA / MEGAIQ [®] parallel	0.97	0.51			
RA / MEGAIQ [®] perpend.	0.93	0.06			
p<0.05*					